

## Catalog Number

49333-EF7

**End Fitting, PEEK, for 7.5 mm ID columns, ARE-Applied Research brand. 1 EA.**



### Additional Info

This End Fitting is made for 4.6mm ID All-PEEK Columns with standard HPLC column 10-32 (tubing inlet) screw threads and 5/8 inch "column tube" screw threads.

Designed for use only with our All-PEEK HPLC Column Hardware and not recommended for use with stainless steel columns.

*Machined from PEEK which ensures a biocompatibility from the beginning of the packing and column use.*

### MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

### Price

**\$57.78**

Printed On 2024/06/01