

## Catalog Number

47450-05

**Adapter**, for use with Low to Medium Pressure HPLC. 10-32 female to female connection, 0.25 mm bore size. PEEK. ARE-Applied Research brand. 1 EA.



### Additional Info

This End Fitting is made for 2.1mm ID PEEK Columns with standard HPLC column 10-32 (tubing inlet) screw threads and 1/4"-28 "column tube" screw threads. The through hole is 0.25mm.

Designed for use only with our All-PEEK HPLC Column Hardware and not recommended for use with stainless steel columns.

*Machined from PEEK which ensures a biocompatibility from the beginning of the packing and column use.*

### MicroSolv Technology Corporation

9158 Industrial Blvd. NE, Leland, NC 28451

tel. (732) 380-8900, fax (910) 769-9435

Email: [customers@mtc-usa.com](mailto:customers@mtc-usa.com)

Website: [www.mtc-usa.com](http://www.mtc-usa.com)

### Price

**\$0.00**

Printed On 2024/05/19